



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319

E mail: procurement@wbmsc.gov.in / mp@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENTS UNDER QUALITY ASSURANCE PROGRAMME FOR THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVT. OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSC/NIT-15/2019

Dated- 28.01.2019

Amendment-I

TECHNICAL SPECIFICATION

Cardiotocography (CTG) equipment

1. Cardiotocography equipment with auto-correlation.
2. Wave form display and digits simultaneously
3. The probe should be waterproof.
4. Vibroacoustic stimulation and twin monitoring facilities desirable.
5. Computer compatibility software facilities.
6. Maturity sensed and automatic foetal movement detection.
7. Alert alarm for tachycardia and Bradycardia.
8. Battery (Lithium) back up for at least 3 - 4 hrs.
9. Patients data storage for at least 10 hrs
10. Should have facility for 3 displays [2 display for FHR (TWIN) and 1 for Uterine pressure]

11. Should be supplied with event marker facility for each foetal movement perceived & noted by the mother.
12. Should be supplied with a Trolley (SS304)
13. Should be supplied with a **inbuilt Thermal** Printer
14. Screen size minimum 5.5"
15. CE & US FDA certification.
16. Should supply **50 rolls** each of CTG thermal paper
17. Conducting jelly / paste - **05 nos.**
18. Continuous supply of paper roll.
19. The rate of paper roll and other consumables should be quoted in Form 9 (a)