



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

**Procurement of Medical Items for the Blood Centres at Hospitals and Medical
Colleges of Govt. of West Bengal**
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-193/2024

Dated-11.03.2024

Amendment-II

REVISED TECHNICAL SPECIFICATION

ITEM-19

REFRIGERATOR FOR KITS & REAGENT STORAGE

WITH DIGITAL DISPLAY

1. Description: Vertical, frost free, CFC free, Single Glass Door, Door with key lock
2. Capacity: From 400 – 600 L
3. Temperature range: Should be from 2⁰C to 8⁰C throughout the Chamber
4. Temperature should be controlled by micro-processor controller
5. Should have digital display of temperature with alarm device (audio-visual)
6. Number of Shelves: Should have minimum 4 nos. with adequate space along with 4 additional perforated plastic tray for air circulation as accessories.
7. Shelves should be adjustable / removable
8. Should be supplied with suitable voltage stabilizer
9. Power supply: 230 V AC, 50-60 Hz

10. Warranty: 2 years;

11. Product Certification: CE / US FDA / BIS certified

****Note Applicable for item no. 1 to 22:**

- 1. Installation Quality (IQ), Operational Quality (OQ), Performance Quality (PQ) certificate to be submitted during installation of the machine.**
- 2. 3^d party calibration certification should be also be provided during installation**