



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

**Procurement of Medical Items for the Blood Centres at Hospitals and Medical
Colleges of Govt. of West Bengal**
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-193/2024

Dated-11.03.2024

Amendment-III

REVISED TECHNICAL SPECIFICATION

ITEM-5

TABLE TOP TUBE SEALER

Feature	Specification
Power supply input frequency	100V AC to 240V AC 50Hz/60Hz
Classification	Protection against electric shock
Operation mode	Once it is turned on , the device is ready to use
RF output power (effective)	20W
Indicator lights / display	Cover open(Yellow), Power(green), Ready(green), Seal(red) or LCD display
Tube detection	Automatic
RF output frequency	40.68 Mhz
Tube specification	Polyvinyl chloride (PVC) tubing up to 6mm (outer diameter)
Sealing time	1 to 3 seconds, adjustable
Dimensions(W*d*h)	155mm x 330mm x 160mm

Feature	Specification
Weight	5.25 kg
Power consumption	Operationg:185W Standby:20W
Fuse	250V,3.15A,type T, slow blow
Warranty	2 years
Product certification:	CE / US FDA / BIS certified