

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V Kolkata-700091

Phone No (033) 40340307/319 E mail: procurement@wbmsc.gov.in

Procurement of Paediatric Cystoscope Set for Department of Paediatric Surgery at Medical College & Hospital, Kolkata

(Submission of Bid through online)

Bid Reference No.: WBMSCL/NIT-850/2025

Dated-15.10.2025

2nd call of Bid Reference Number: WBMSCL /NIT-650/2025;dated –28.07.2025

The following amendment has been made in the tender document,

Amendment – 1

REVISED TECHNICAL SPECIFICATION

Paediatric Cystoscope Set

- 1. Cystoscope Sheath, (9-10) Fr. with (4-5) Fr. working channel, length (14-16) cm
- 2. Obturator compatible with cystoscope sheath (9-10) Fr.
- 3. Forceps for foreign body removal, Double action jaw, 3 Fr., Length (23-28) cm Compatible with paediatric cystoscope 9 Fr.
- 4. 4 nos. Bugbee electrode, rigid, (2-5) Fr., (53-60) cm Compatible with paediatric cystoscope 9 Fr.
- 5. Paediatric working element and compatible zero degree miniature straight forward telescope zero degree, Diameter (1.2 -1.9) mm, Length 20cm
- 6. Paediatric Resectoscope sheath 9Fr., Compatible with Paediatric working element
- 7. Obturator compatible with Resectoscope sheath 9 Fr.
- 8. Telescope Bridge compatible with paediatric Resectoscope sheath
- 9. Coagulating electrode hook shaped without ball (6 units) and compatible with paediatric working element
- 10. **Certificate**: The bidder/OEM should submit valid CDSCO Certificate/ Registration/ License for both the manufacturer(s) and importer(s) as applicable.
- 11. All the products should be from one manufacturer.