



## **Notice Inviting e-Tender**

**West Bengal Medical Services Corporation Limited**

**Swasthya Sathi**

**GN-29, Salt Lake, Sector-V**

**Kolkata-700091**

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

**Supply and Commissioning of 02(Two) units of Heater Cooler machine for the Department of C.T.V.S, IPGME&R-SSKM Hospital, Kolkata**

(Submission of Bid through *online*)

(2<sup>nd</sup> call of WBMSCL/NIT-953/2025, Dated-24.11.2025)

**Bid Reference No.: WBMSCL/NIT-134/2026**

**Dated-12.02.2026**

### **AMENDMENT-I**

## **Revised Technical Specification**

### **Heater Cooler Machine**

1. The unit shall be capable of supplying water of temperature 2 - 41 degree Celsius for patient circuit. 2-10 degree Celsius for Cold cardioplegia and 15-41 degree Celsius for warm cardioplegia circuit.
2. The unit should have 3 independent tanks and 3 separate circuits, and these circuits should be able to control patient's temperature and also heating and cooling of cardioplegia and should work simultaneously. Should have easy de-airing process. Should have easy drainage system. Should be supplied with bacteria collection aerosol canisters.
3. The accuracy should be at least 0.1C. Settings should be adjustable to approx. 0.1
4. The heater cooler unit should also be compatible to get integrated into the heart lung machine.
5. **Maintenance Service During Warranty Period**
  - During the warranty period supplier must ensure two planned preventive maintenance (PPM) in a year along with corrective/breakdown maintenance whenever required.
6. The bidder/OEM should submit valid CDSCO Certificate/Registration/License for both the manufacturer(s) and importer(s) as applicable.