



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and commissioning of various ophthalmic equipment across different Health Facilities of the Govt of West Bengal under the project NPCB and VI

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-434/2025

Dated-24.05.2025

2nd call for items at Schedule No .V, VI, VII, VIII, IX, XI,XII and XIII of
Bid Reference Number WBMSCL/NIT-293/2025 Date-16.04.2025

Amendment I

Revised Technical Specification

Schedule No.VIII Laminar Flow Hood

1. Size: 4 x 2 x 2 (ft)
2. Inner and outer chamber: SS304
3. Air flow: Vertical
4. Air Velocity: between 0.3 to 0.6 m/s $\pm 20\%$
5. HEPA Efficiency Pre-filter – 99.997% at 0.3 microns with DOP test Washable with an arrestance up to 90% at 5 microns

6. Sound Level – Less than 65 dB (Acoustic)
7. UV light source and visual light source Light
8. Intensity (visual): Lux 800 to 1300 Lux
9. Federal Power Supply: 220 V Single phase 50 Hz
10. 3rd party Calibration required during DOP test & particle test during the time of installation.
11. **Safety & Standard:** Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable