



Notice Inviting e-Tender

**West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091**

Phone No (033) 40340307/319
E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Medical Equipment for Physical Medicine & Occupational Therapy Equipment at IPGME&R and SSKM Hospital of the Govt. of West Bengal.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-139/2026

Dated-12.02.2026

3rd call bid reference no. WBMSCL/NIT-478/2023 Dated-14.08.2023

**The following amendment has been made in the tender document,
Amendment – I (Schedule-I)**

REVISED TECHNICAL SPECIFICATION

ELECTRIC TILT TABLE

1. The unit should have electric height and tilting operation with hand switch control.
2. The unit should have safe working load and lifting capacity upto 225 kg.
3. The unit should have height adjustment from 47 to 100 cm.
4. The unit should have adjustable angle dual foot boards with positive and negative (+ 15° to -30°).
5. The unit should have adjustable angle backrest facility with anti trap design (0 to +75°).
6. The unit should have tilting upto 90° from horizontal.
7. The unit should have 195 cm – 200 cm X 68 cm – 70 cm (Length x Width)
8. The unit should have electric height and tilting operation with hand switch control.

9. The unit should have large wheel design, with central locking & steering facility.
10. The unit should have fitted Tilt angle inclinometer as standard.
11. The unit should have lowers to wheelchair height for ease of patient transfer.
12. The unit should have split leg divided into 2 parts with locking facility.
13. The unit should be supplied with patient worktable, handgrips, Set of 3 nos. Harnesses, battery backup, hip-leg stabilizers, and head support cushions of same manufacturer.
14. The bidder/ OEM should have valid CDSCO Certificate/Registration/License for both the manufacturer(s) and importer(s) as applicable.
15. It works on 220-230V AC/50 Hz.